



# UNITED STATES PATENT AND TRADEMARK OFFICE

4/17  
UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER OF PATENTS AND TRADEMARKS  
Washington, D.C. 20231  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/732,200	12/07/2000	Volker Rasche	PHD 99,179	9483

7590

08/23/2002

Jack E. Haken  
Corporate Patent Counsel  
U.S. Philips Corporation  
580 White Plains Road  
Tarrytown, NY 10591

EXAMINER

KAO, CHIH CHENG G

ART UNIT

PAPER NUMBER

2882

DATE MAILED: 08/23/2002

Please find below and/or attached an Office communication concerning this application or proceeding.

## Office Action Summary

Application No.

09/732,200

Applicant(s)

RASCHE ET AL.

Examiner

Chih-Cheng Glen Kao

Art Unit

2882

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133).
- Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☐ Responsive to communication(s) filed on \_\_\_\_.
- 2a) ☐ This action is FINAL. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-20 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-20 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 07 December 2000 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- 11) ☐ The proposed drawing correction filed on \_\_\_\_ is: a) ☐ approved b) ☐ disapproved by the Examiner.  
If approved, corrected drawings are required in reply to this Office action.
- 12) ☐ The oath or declaration is objected to by the Examiner.

### Priority under 35 U.S.C. §§ 119 and 120

- 13) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).  
a) ☒ All b) ☐ Some \* c) ☐ None of:  
1. ☒ Certified copies of the priority documents have been received.  
2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_.  
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).  
\* See the attached detailed Office action for a list of the certified copies not received.
- 14) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. § 119(e) (to a provisional application).  
a) ☐ The translation of the foreign language provisional application has been received.
- 15) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. §§ 120 and/or 121.

### Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892) 4) ☐ Interview Summary (PTO-413) Paper No(s). \_\_\_\_
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948) 5) ☐ Notice of Informal Patent Application (PTO-152)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449) Paper No(s) \_\_\_\_ 6) ☐ Other: \_\_\_\_

## DETAILED ACTION

### *Claim Rejections - 35 USC § 103*

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

1. Claims 1-4, 7, 11, 17, 18, and 20 are rejected under 35 U.S.C. 103(a) as being unpatentable over Rattner (US Patent 6,213,638) in view of Kresse (US Patent 4894855).

Rattner discloses an x-ray device with a source (Fig. 1, #2) and detector (Fig. 1, #3) mounted at an invariable distance on a rigid common holding device (Fig. 1, #1) changed in a plane defined by the supporting members (Fig. 1, #1, b), connected to a supporting or displacement device (Fig. 1, #5) composed of a plurality of hinged, serially interconnected supporting members (Fig. 1, #7) as a robot arm to position completely (Fig. 1, "b", "α", and "β") along 6 axes. The individual supporting members can be individually controlled (Fig. 1, "g" represented by individual motors), while the holding device in the form of a C-arm (Fig. 1) is connected to the holding device by way of a hinge (Fig. 1, #4 and "β"), which rotates about a horizontal axis of the hinge (Fig. 1, "A").

However, Rattner does not seem to specifically disclose plane hinges.

Kresse teaches plane hinges (col. 2, lines 61-68, and col. 3, lines 1-2).

It would have been obvious, to one having ordinary skill in the art at the time the invention was made, to have the plane hinges to move the holding device of Kresse with the

Art Unit: 2882

device of Rattner, since one would be motivated to use the hinges to create free accessibility or positioning of the c-arm to the patient when operating the x-ray device as implied from Kresse (col. 1, lines 60-68, and Figure).

2. Claim 5 is rejected under 35 U.S.C. 103(a) as being unpatentable over Rattner in view of Kresse as applied to claim 1 above, and further in view of Hollstein (US Patent 3281598).

Rattner in view of Kresse suggests a device as recited above.

However, Rattner does not seem to specifically disclose a hinge connected to the holding device permitting 360 degree rotation about an axis.

Hollstein teaches a hinge connected to the holding device permitting 360 degree rotation about an axis (Fig. 3 and col. 3, lines 21-32).

It would have been obvious, to one of ordinary skill in the art at the time the invention was made, to have the rotating hinge of Hollstein with the suggested device of Rattner in view of Kresse, since one would be motivated to direct x-rays to all directions as shown by Hollstein (col. 3, lines 25-26) in order to obtain an x-ray image from any direction.

3. Claim 6 is rejected under 35 U.S.C. 103(a) as being unpatentable over Rattner in view of Kresse as applied to claim 1 above, and further in view of Holmström (US Patent 3,784,837).

Rattner in view of Kresse suggests a device as recited above.

However, Rattner does not seem to specifically disclose a holding device composed of at least two holding members for the source and detector.

Art Unit: 2882

Holmström discloses a holding device composed of at least two holding members for the source and detector (Fig. 1).

It would have been obvious, to one of ordinary skill in the art at the time the invention was made, to prepare the holding device of Holmström with the suggested device of Rattner in view of Kresse, since one would be motivated to have separate holding members to move the x-ray source and detector as freely as possible around the patient as shown by Holmström (col. 1, lines 6-8) and to keep the x-ray source and detector independently controlled for proper alignment (col. 2, lines 1-9) to send x-rays and obtain a signal.

4. Claims 8, 10, 14, and 15 are rejected under 35 U.S.C. 103(a) as being unpatentable over Rattner in view of Kresse as applied to claim 1 above, and further in view of Travanty et al. (US Patent 4,987,583).

Rattner in view of Kresse suggests a device as recited above.

However, Rattner does not seem to specifically disclose a mechanical contact sensor producing a signal to monitor distance between the examined object and the source or detector with braking when a threshold when the distance is below a threshold.

Travanty et al. discloses a mechanical contact sensor producing a signal (col. 3, lines 63-66) to monitor distance between the examined object and the source or detector (abstract, lines 2-4 and col. 3, lines 50-56) with braking when a threshold when the distance is below a threshold (col. 2, lines 11-14).

It would have been obvious, to one of ordinary skill in the art at the time the invention was made, to prepare the mechanical contact sensors producing a signal and braking of Travanty

Art Unit: 2882

et al. with the suggested device of Rattner in view of Kresse, since one would be motivated to incorporate all those things to protect the examined object or patient from being severely hurt by contact with the source or detector as shown by Travanty et al. (col. 1, lines 38-42, col. 2, lines 11-14).

5. Claims 9 and 12 are rejected under 35 U.S.C. 103(a) as being unpatentable over Rattner in view of Kresse and Travanty et al. as applied to claim 8 above, and further in view of Hinton et al. (US Patent 5485502).

Rattner in view of Kresse suggests a device as recited above.

However, Rattner does not seem to specifically disclose software control of the c-arm and ultrasound monitoring of the object and x-ray device.

Hinton et al. teaches software control of the c-arm (col. 5, lines 14-19) and ultrasound monitoring of the object and x-ray device (Abstract, lines 1-3, col. 1, lines 48-53, and col. 12, lines 53-58).

It would have been obvious, to one of ordinary skill in the art at the time the invention was made, to have the software control of Hinton et al. with the suggested device of Rattner in view of Kresse and Travanty et al., since one would be motivated to have use a computer and software to provide better control of the motion of c-arm so as to follow an efficient path between two positions and to avoid collision between the various elements of that system as shown by Hinton et al. (col. 2, lines 10-15) in order to save time and increase safety.

It would have been obvious, to one of ordinary skill in the art at the time the invention was made, to have ultrasonic monitoring of Hinton et al. with the suggested device of Rattner in

Art Unit: 2882

view of Kresse and Travanty et al., since one would be motivated to have use the sensors to avoid collision between the various elements of that system as shown by Hinton et al. (col. 2, lines 10-15, and col. 12, lines 47-53) in order to increase safety.

6. Claim 13 is rejected under 35 U.S.C. 103(a) as being unpatentable over Rattner in view of Kresse and Holmström as applied to claim 6 above, and further in view of Yamamoto (JP 06-105831).

Rattner in view of Kresse and Holmström suggests a device as recited above.

However, Rattner does not seem to specifically disclose the distance between source and detector, which can change.

Yamamoto teaches the distance between source and detector which can change (Paragraph [0005], and Drawing 2, "E-F").

It would have been obvious, to one of ordinary skill in the art at the time the invention was made, to have changing distances between the source and detector of Yamamoto with the suggested device of Rattner in view of Kresse and Holmström, since one would be motivated to move the detector to the patient as close possible without hitting the applicant as shown by Yamamoto (Paragraph [0005]) for safety purposes.

7. Claim 16 is rejected under 35 U.S.C. 103(a) as being unpatentable over Rattner in view of Kresse and Travanty et al. as applied to claim 8 above, and further in view of Ninomiya et al. (JP 11-285492).

Rattner in view of Kresse and Travanty et al. suggests a device as recited above.

Art Unit: 2882

However, Rattner does not seem to specifically disclose a separate video system to monitor the motion of the c-arm.

Ninomiya et al. teaches a separate video system to monitor the motion of the c-arm (Abstract, Problem to be Solved).

It would have been obvious, to one of ordinary skill in the art at the time the invention was made, to have the separate video system of Ninomiya et al. with the suggested device of Rattner in view of Kresse and Tranvanty et al., since one would be motivated to keep track of the movement safely and reliably when they are operated as shown by Ninomiya et al. (Abstract, Problem to be Solved).

8. Claim 19 is rejected under 35 U.S.C. 103(a) as being unpatentable over Rattner in view of Kresse as applied to claim 1 above, and further in view of Stivender et al. (US Patent 4358856).

Rattner in view of Kresse suggests a device as recited above.

However, Rattner does not seem to specifically disclose a rotating connection point to the room.

Stivender et al. teaches a rotating connection point to the room (Abstract, lines 1-3, and Fig. 2, #13-15)

It would have been obvious, to one of ordinary skill in the art at the time the invention was made, to have the rotating connection point of Stivender et al. with the suggested device of Rattner in view of Kresse, since one would be motivated to include another rotation of axis in a multiaxial apparatus to be able to do substantially all examination procedures while keeping the



Art Unit: 2882

patient remaining at a constant level in coincidence with an isocenter as implied from Stivender et al. (col. 4, lines 1-5), so the patient does not have to be moved or disturbed.

***Response to Arguments***

9. Applicant's arguments with respect to claims 1-20 have been considered but are moot in view of the new ground(s) of rejection.

***Conclusion***


Any inquiry concerning this communication or earlier communications from the examiner should be directed to Chih-Cheng Glen Kao whose telephone number is (703) 605-5298. The examiner can normally be reached on M - Th (8 am to 5 pm).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Robert Kim can be reached on (703) 305-3492. The fax phone numbers for the organization where this application or proceeding is assigned are (703) 308-7722 for regular communications and (703) 308-7724 for After Final communications.

Any inquiry of a general nature or relating to the status of this application or proceeding should be directed to the receptionist whose telephone number is (703) 308-0956.



gk  
August 21, 2002



ROBERT H. KIM  
SUPERVISOR, PATENT EXAMINER  
BIOLOGY CENTER 2800